

**United States Department of Commerce  
National Institute of Standards and Technology**

**~ Certificate of Measurement Traceability ~  
for**

**Kentucky**

**a participant in the NIST Weights and Measures  
Measurement Assurance Program for the Year(s)**

**2006**

**The State standards are traceable to NIST and the laboratory  
has *conditionally* demonstrated\* measurement proficiency  
through training and interlaboratory comparisons.**



*Georgia L. Harris*  
for the National Institute of Standards and Technology

Laboratory Metrology, Group Leader, Weights and Measures Division  
Title

December 29, 2005  
Effective Date

**\*Demonstrated  
Areas of Proficiency**

**Mass**

Echelon III  
2 kg - 1 mg  
1000 lb - 0.001 lb  
8 oz - 0.03125 oz

**Volume**

Volume transfer  
1000 gal - 5 gal

*Conditional status: Updates to  
the quality manual are needed*

